



REPLACEMENT SHEET DRAWING

	AsiaGe MT Detection Kit		
	+	-	
Standard*	32	0	32
Standard	0	73	73
	32	73	105

Sensitivity = 100%

Specificity = 100%

PPV = 100% NPV = 100%

Smear
~~Smear~~⁺, Culture⁺, Nested PCR⁺
 PPV: Positive Prediction
 NPV: Negative Prediction Value

FIG. 5